E. I. du Pont de Nemours and Company
Wilmington, Delaware 19880-0024

Certificate of Sterility - Gamma Irradiation

Sterility Run Number:	**M1301626B**
Product Release Date:	**10/04/2013 03:10:02**
Minimum Specified Dose (kGy):	**31.0**
Maximum Specified Dose (kGy):	**55.0**
Minimum Validation Dose (kGy):	**30.0**
Minimum Delivered Dose (kGy):	**33.7**
Maximum Delivered Dose (kGy):	**49.7**

Product Number:	**IC253BWH2X0025CS**
Product Name:	**DUPONT(TM) TYVEK® ISOCLEAN® COVERALL**
Product Size:	**2X**
Manufacturing Lot Number:	**AHS1339114**
Sterilization Lot Number:	**AHS1339114**

SAMPLE

DuPont certifies the product number and lot numbers above have been processed and released according to DuPont specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^-6. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved on-line at www.personalprotection.dupont.com or contact DuPont Customer Service at 1-800-931-3456.

Approved By: ***Sandy S. Byrd***
Quality Assurance

E. I. du Pont de Nemours and Company
This report is provided at no charge solely to assist DuPont customers in meeting their quality systems needs. The recipient of this report agrees to evaluate such information, to make an independent decision regarding the suitability of such information for its application, and only use such information pursuant to then current good product stewardship principles and all regulatory requirements applicable to your business.